

## Challenges with Life Surface Imaging

Lifong ZOU<sup>1\*</sup>, Robin RICHARD<sup>2</sup>, John BLYTHE<sup>3</sup>, Nikolaos DONOS<sup>4</sup>

<sup>1</sup> The Bioengineering Centre, Bart's and The London School of Medicine and Dentistry, Queen Mary University of London, UK;

<sup>2</sup> Cavendish Implants Ltd, UK; <sup>3</sup> Bart's and The London Dental Hospital, NHS Trust, UK;

<sup>4</sup> Centre for Oral Clinical Research and Centre for Immuno-Biology & Regenerative Medicine, Bart's and The London School of Medicine and Dentistry, Queen Mary University of London, UK

DOI: 10.15221/18.064 <http://dx.doi.org/10.15221/18.064>

### Abstract

Stereo photogrammetry has a clear role in clinical research, treatment planning and effectiveness assessment within the domain of medicine and dentistry. Recently more computational automatic 3D facial analysis methodology has emerged to increase the efficiency for data processing, this is particularly beneficial for clinical trials as it often request large data size. The quality of automation is depend on two aspects, the algorithm/program need to be robust enough to cope the variance from the individuals and scenarios, and the raw data capturing need to be accurate. To the latter aspect, despite practical guide has published on top of the manufacturer suggestions, a detailed understanding of the error causation are requested, to eliminate further to the uncertainty from the process of facial surface capturing. This paper presents challenges of life surface capturing due to its dynamic nature and proposed a possible and practical solution.

**Keywords:** 3D facial imaging, 3D facial scan, 3D facial measurements, life surface capturing, life surface dynamics

### Introduction

Stereo photogrammetry has a clear role in treatment decision and planning [1]; It is commonly used in maxillofacial surgery treatment plan (T.J Maal et al 2011, Van Loon et al 2010); 3D facial asymmetry analysis provided the soft tissue deficit distribution and magnitude prior to the facial reconstructive operation to head and neck cancer patients (Kansy K et al 2018); facilitating treatment plan for Cleft patients (Zhang C et al 2018); management decision based on quantitative facial image analysis of patients with facial dysmorphism and hypertrophic muscles of mastication (LHH Cheng et al, 2011, L Collier et al 2009). Moreover, Craniofacial researchers make heavy use of the facial morphometric analyses (Metzler P et al, 2012, Toma AM et al, 2008, Kau CH et al 2007, Aldridge K et al 2005, Moss JP et al 2003, Altobelli DE et al 1993), and Genome association to the facial morphology has its increased attention in recent years (Kraemer M, et al 2018; Shaffer JR et al, 2016; Boehringer S et al, 2011).

Accuracy and reproducibility of 3D facial scanning systems were investigated by number of researchers (Lubbers HT et al 2010, Heike CL et al 2009, Toma AM et al 2009, Weinberg SM et al 2006, Aldridge et al, 2005). 3dMDface is one of the favourable 3D facial surface capturing device, due to its fast speed of capturing (0.25s) and the linear accuracy of 0.2mm.

A practical guide to facial image acquisition was published (Heike CL, 2010), that reviewed the common issues affect the accuracy of the data capturing and solutions to achieve optimal performance on top of the manufacturer's recommendation in practice. Despite the elimination was made in possible movements from body and face, policy of avoiding the artefacts as well as the environmental control. The dynamics of the life surface are inevitable and it need to be aware and considered properly for facial image analysis and data interpretation for an objective quantitative information.

### The dynamics of life facial surface

A temporal facial surface capturing device 3dMDface.t system was used to capture a sequence of 3D facial surfaces at the rate of 6 frames per second at high resolution for 20 seconds, this permit the selection of appropriate or specified posture or moment of facial image; The sittings of the device was completely satisfied the requirements from the manufacturer; an adjustable chair for patient to sit, which can be moved freely to justify the distance from patient to the camera units and height according to the size of the face and height of the body from each individuals in relation to the device; the same set of instructions were given to every individuals for a neutral body and head posture after the justification of the height and distance to be within the depth of view to the two modular camera units; a back teeth occlusion lightly applied.

Subjects were NHS patients who participated the clinical trial (COCR00013) or NHS patients referred by maxillofacial consultant; during the facial image analysis process, four cases were identified as examples for the presentation of the dynamics of life surface that are inevitably adhered (1) breathing, (2) emotion, (3) occlusion and (4) facial hair.

### Four adhered causation for uncertainty

- (1) The inhaling and exhaling motion brought depth difference of 0.07mm at most facial areas and 1.34mm at the left and right alar as well as upper and lower lips (Fig.1 Left), while the images between the breaths for 3 seconds the depth difference can be reduced to 0.01mm over the whole face except eye lids (Fig.1 Right).





Fig.1 Demonstrates the effect of breathing on the left image with 0.07 mm at the cheek and 1.34mm at the alar; When between the breathe the depth difference reduced to 0.01mm

- (2) To give the same instructions to one subject at the same visit and take two sequences of facial scans, select one image from each sequence, a slight emotional differences were captured and expressed by a concentrated differences around two corners of the mouth, to be at the negative difference of 1.48mm, and penetrated to the cheek at the positive difference of 0.32mm (Fig.2).




Fig.2 Demonstrates the effect of emotion changes of 0.32mm at cheek and 1.48 mm at corner of the mouth

- (3) Despite an instruction of light occlusion at back teeth was give, the facial difference purely caused by the mandibular position was found from two images of one captured sequence, to be as large as 1.04mm at middle of the chin and 0.02mm at cheek where showed no effect (Fig 3).



*Fig.3 Demonstrates the effect of mandible positioning at 1.04 mm depth difference in the middle of chin while the cheek area has no effect with difference of only 0.02 mm*

- (4) Facial hair is an obvious factor which can be easily controlled theoretically, however it is still appeared to be one of the often occurred error sources, can be as large as 1.41mm Fig.4.



*Fig.4 Demonstrates the effect of facial hair at 1.41mm depth difference*

### One approach to the solution

Four cases demonstrated four random possible error sources when facial image was captured, due to its uncontrollable nature, particularly to the clinical trial that requests multiple consecutive facial image captures at difference health and emotional conditions, different days and times. To eliminate the random errors as shown above, image averaging technique can be one effective approach. Particularly, if the averaging process can be based on an automatic marking process.

The facial averaging process involved with removal of translation, rotation and size differences, and the method of averaging is made along the surface norm. The initial step is based upon a set of critical markers on every 3D facial images [24].

The facial landmarks may be used to warp the template mesh to each individual face using a method known as the thin plate spline warp, to achieve close alignment of the template mesh to the subject's face that matches exactly at the landmarks, and approximately elsewhere. The vertices of the template mesh are then projected onto the subject face by finding the closest point on the subject face to the given vertex. Finally, a smoothing procedure is applied to vertices of the template mesh to ensure an even distribution of vertices on the subject face, excluding those directly adjacent to a landmark and those on the boundary of the template mesh. The smoothing procedure consists of a Laplacian smoothing operation, followed by projection to the closest point on the subject face and repeated until the maximum change of combined smoothing plus projection operation is below a pre-defined threshold.

The manual annotation of landmarks is a known source of variance and it is labour intensive. Siang S et al [25] introduced an improved method for automatically detect landmarks on 3D human face data by three steps: First, geometric information was used to locate 17 prominent points. Then a deformable transformation between target mesh and data mesh determined 20 established landmarks and located them more accurately than with the geometric method alone. This method has an average error of 2.64 mm over a sample of 115 heads.

Further studies need to implement this proposal and to conduct a though evaluation before it applied to clinical studies.

## References:

- [1] Laura Verzé, Francesca Antonella, Bianchi and Guglielmo Ramieri (2014), Three-dimensional laser scanner evaluation of facial soft tissue changes after LeFort I advancement and rhinoplasty surgery: patients with cleft lip and palate vs patients with nonclefted maxillary retrognathic dysplasia (control group), *Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology*; Vol:117-4, Pg:416-23;
- [2] Joanneke M. Plooi, Thomas J. J. Maal, Piet Haers, Wilfred A. Borstlap, Anne Marie Kuijpers-Jagtman, Stefaan J. Berge (2011), Digital three-dimensional image fusion processes for planning and evaluating orthodontics and orthognathic surgery. A systematic review; *International Journal of Oral and Maxillofacial Surgery* Vol:40-4, Pg:341-52;
- [3] Thomas J.J.Maal, Bram van Loon, Joanneke M. Plooi, Frits Rangel, Anke M. Ettema, Wilfred A.Borstlap, and Stefaan J. Bergé (2010), Registration of 3-Dimensional Facial Photographs for Clinical Use, *Journal of Oral and Maxillofacial Surgery*, Vol:68-10, Pg:2391-2401
- [4] Kansy K, Hoffmann J, Alhalabi O, Mistele N, Freier K, Mertens C, Freudisperger C and Engel M (2018), Subjective and objective appearance of head and neck cancer patients following microsurgical reconstruction and associated quality of life—A cross-sectional study; *Journal of Cranio-Maxillofacial Surgery*, Vol:46-8, Pg:1275-84;
- [5] Zhang C, Miller SF, Roosenboom J, Wehby GL, Moreno Uribe LM, Hecht JT, Deleyiannis FWB, Christensen K, Marazita ML, Weinberg SM (2018); Soft tissue nasal asymmetry as an indicator of orofacial cleft predisposition, *Am J Med Genet A*. 2018 Jun;176(6), Pg:1296-1303. doi: 10.1002/ajmg.a.38688;
- [6] VT Nguyen,T Nguyen and TJagomägi (2018), Nasolabial aesthetics of patients with repaired unilateral cleft lip and palate: A comparison of three rating methods in two countries, *Journal of Cranio-Maxillofacial Surgery* Vol:46-8, Pg1385-9;
- [7] LHH Cheng, L Zou, R Bhandari, N Ali, J Collier, and B Lee (2011); Management decision based on quantitative facial laser scanning of patients with facial dysmorphophobia and hypertrophic muscles of mastication; Abstracts of BAOMS annual scientific meeting; Nice, France, 22-24th June 2011; Vol49, Pg:41.
- [8] J. Collier and L Zou (2009); Quantitative measurement of the response of masseteric hypertrophy to treatment with Botulinum Toxin Type-A; *British Journal of Oral and Maxillofacial Surgery* 47:e1-e34.
- [9] Philipp Metzler, Yi Sun, Wolfgang Zemann, Alexander Bartella, Marc Lehner, Joachim Anton Obwegeser, Astrid L. Kruse-Gujer and Heinz-Theo Lübbers (2014), Validity of the 3D VECTRA photogrammetric surface imaging system for cranio-maxillofacial anthropometric measurements, *Oral and Maxillofacial Surgery*, Vol:18-3, Pg 297–304;
- [10] Val Lambros and Gideon Amos (2016), Three-Dimensional Facial Averaging: A Tool for Understanding Facial Aging Society of Plastic Surgeons Volume 138, Number 6, Ideas and Innovations;
- [11] Toma AM, Zhurov A, Playle R, et al. A three-dimensional look for facial differences between males and females in a British-Caucasian sample aged 151/2 years old. *Orthod Craniofac Res* 2008; Vol:11, Pg:180-185;
- [12] Kau CH, Richmond S and Incrapera A (2007) Three-dimensional surface acquisition systems for the study of facial morphology and their application to maxillofacial surgery. *Int J Med Robot* 2007 Vol:3, Pg:97-110

- [13] Aldridge K, Boyadjiev SA and Capone GT (2005), Precision and error of three-dimensional phenotypic measures acquired from 3dMD photogrammetric images. *Am J Med Genet, Part A*; Vol:138A-3, Pg:247-253,
- [14] Moss JP, Ismail SF, Hennessy RJ. Three-dimensional assessment of treatment outcomes on the face. *Orthod Craniofac Res* 2003;Vol:6, Pg:126-131;
- [15] Kraemer M, Huynh QB, Wieczorek D, Balliu B, Mikat B and Boehringer S (2018); Distinctive facial features in idiopathic Moyamoya disease in Caucasians: a first systematic analysis, *PeerJ*. 2018 Jun 27;6:e4740. doi: 10.7717/peerj.4740. eCollection 2018;
- [16] Shaffer JR, Orlova E, Lee MK, Leslie EJ, Raffensperger ZD, Heike CL, Cunningham ML, Hecht JT, Kau CH, Nidey NL, Moreno LM, Wehby GL, Murray JC, Laurie CA, Laurie CC, Cole J, Ferrara T, Santorico S, Klein O, Mio W, Feingold E, Hallgrimsson B, Spritz RA, Marazita ML, Weinberg SM (2016); Genome-Wide Association Study Reveals Multiple Loci Influencing Normal Human Facial Morphology. *PLoS Genet*. 2016 Aug 25;12(8):e1006149. doi: 10.1371/journal.pgen.1006149. eCollection 2016 Aug;
- [17] Boehringer S, van der Lijn F, Liu F, Günther M, Sinigerova S, Nowak S, Ludwig KU, Herberz R, Klein S, Hofman A, Uitterlinden AG, Niessen WJ, Breteler MM, van der Lugt A, Würtz RP, Nöthen MM, Horsthemke B, [18] Wieczorek D, Mangold E, Kayser M (2011), Genetic determination of human facial morphology: links between cleft-lips and normal variation, *Eur J Hum Genet*. 2011 Nov;19(11):1192-7. doi: 10.1038/ejhg.2011.110. Epub 2011 Jun 22;
- [18] Lubbers HT, Medinger L, Kruse A, et al. Precision and accuracy of the 3dMD photogrammetric system in craniomaxillofacial application. *J Craniofac Surg* 2010, Vol:21, Pg:763-767
- [19] Heike CL, Cunningham ML, Hing AV, et al. Picture perfect? Reliability of craniofacial anthropometry using three-dimensional digital stereophotogrammetry. *Plast Reconstr Surg* 2009; Vol:124, Pg:1261-1272
- [20] Toma AM, Zhurov A, Playle R (2009), Reproducibility of facial soft tissue landmarks on 3D laser-scanned facial images. *Orthod Craniofac Res* 2009;Vol:12, Pg:33-42;
- [21] Weinberg SM, Naidoo S, Govier DP (2006) Anthropometric precision and accuracy of digital three-dimensional photogrammetry: comparing the Genex and 3dMD imaging systems with one another and with direct anthropometry. *J Craniofac Surg* 2006;Vol:17, Pg:477-483
- [22] Philipp Metzler, Lea S. Bruegger, Astrid L. Kruse Gujer, Felix Matthews, Wolfgang Zemmann, Klaus W. Graetz and Heinz-Theo Luebbbers(2012), Craniofacial Landmarks in Young Children: How Reliable Are Measurements Based on 3-Dimensional Imaging? *The Journal of Craniofacial Surgery* & Volume 23, Number 6, November 2012
- [23] Carrie L Heike, Kristen Upson, Erik Stuhaug and Seth M Weinberg (2010), 3D digital stereophotogrammetry: a practical guide to facial image acquisition, *Head & Face Medicine* 2010, 6:18 <http://www.head-face-med.com/content/6/1/18>
- [24] Shaweesh AI , Thomas CD , Bankier A , Clement JG (2004), Delineation of facial archetypes by 3d averaging, *Annals of the Royal Australasian College of Dental Surgeons* [01 Oct 2004, 17:73-79];
- [25] Liang S, Wu J, Weinberg SM and Shaprio LG (2013); Improved detection of landmarks on 3D human face data; *Conf Proc IEEE Eng Med Biol Soc* 2013:6482-5, doi:10.1109/EMGC:2013:6611039;